

SODA LIME

CO₂ ABSORBENT



INSTRUCTIONS FOR USE

TRADE NAMES: Sofnolime, Sofnolime Solo, Sodasorb, Sodasorb L F Ch ialime, L imepak, L imedic, Super L imedic, Sodaesia, Tgersorb, Durasorb

⚠ WARNINGS & PRECAUTIONS

Avoid contact with eyes, skin and clothing.

Soda lime carbon dioxide absorbents have only been verified for use in combination with medical procedures and should not be used outside the scope of this instruction for use.

Do not use Soda lime carbon dioxide absorbents with chloroform or trichloroethylene as carbon monoxide will be produced.

Soda lime carbon dioxide absorbents are not an antimicrobiological filter or inhibitor. It is the responsibility of the user to take appropriate measures regarding patient cross contamination.

Soda lime carbon dioxide absorbent must be used with its shelf life. The expiration date is printed on device packaging.

There is potential for dust generation and migration into the breathing circuit due to poor handling, and care should be taken to minimise the risk of dust. Avoid breathing dust, which can cause irritation.

Discard the last 10mm or so of the pack of resealable containers, which may consist of dust or small granules as a result of transportation/handling.

Do not flush dry gas through the absorber when not in use as this can dry out the soda lime leading to a loss of performance.

Ensure storage of the product is conducted in accordance with the instructions.

Re-sealable containers, once opened, must be properly resealed to maintain product integrity and should in all cases be used within one month of opening.

The colour change may reverse if the device is left alone, except for Sodasorb L F which has a permanent colour change.

The absorbent may contain traces of different soda lime grades of a different colour.

GENERAL DESCRIPTION

Soda lime CO₂ absorbents are intended for medical purposes, to remove carbon dioxide from gases in breathing circuits, in systems such as anesthesia and respiratory therapy equipment.

INTENDED USE

Soda lime is intended to remove carbon dioxide expired from patients in a medical closed or semi-closed breathing circuit.

INTENDED USER

This device must be used by trained professional users familiar with the operation of closed or semi-closed loop breathing circuits.

INDICATION

The device is indicated whenever carbon dioxide needs to be removed from medical closed or semi-closed breathing circuits.

CONTRAINDICATIONS

This device has no known contraindications.

CLINICAL BENEFIT

This device supports the regulation of patient carbon dioxide levels when used in a closed or semi-closed breathing circuit, through the absorption of carbon dioxide.

LIMITATIONS

Soda lime carbon dioxide absorbents are for use only with Oxygen, Nitrous Oxide, Halothane, Enflurane, Desflurane, Sevoflurane and Isoflurane.

PERFORMANCE

Approximately 150 litres of carbon dioxide are absorbed per kilogram of soda lime.

PREPARATION



Before conducting any preparation activities ensure that the warnings and precautions of this instructions for use have been read and understood.

1. OPENING/UNWRAPPING

- Carefully open container to avoid any spills.

2. FILLING AND CHANGING

- Anaesthesia and respiratory therapy equipment absorbent canisters should be cleaned when the absorbent container is changed, with particular attention paid to any screens as they are susceptible to obstruction.
- Absorbents should always be handled gently to avoid fragmentation and dust formation. When a container is emptied, care should be taken to remove dust particles.
- Loose-fill absorbent should be poured with care into the canister while the canister is rotated, stopping occasionally to tap the sides. The canister should be filled completely but not overfilled. A small space should be left at the top to promote even gas flow through the canister. The upper layer of absorbent should be level.
- For pre-filled carbon dioxide absorbent round cartridge, completely remove any outer wrapping. Inspect the body top and bottom grids for defects, damage or obstructions.

3. PRE-CHECK

- Users are responsible for performing equipment pre-checks in accordance with machine instructions.
- The device must be checked for damage prior to use.

USE

- Before use, check the breathing circuit for gas flow restrictions.
- The Soda lime carbon dioxide absorbents contain an indicator that turns from white to purple or from pink to white in use. The absorption of carbon dioxide is indicated by a gradual colour change in the direction of air flow through the absorbent. The intensity of colour change may vary from one application to another. Colour change is only a guide, the anaesthetist should rely primarily on carbon dioxide monitors as well as time and volume calculations to determine remaining absorptive capacity.
- System filters are to protect the patient from contaminants and should be fitted in accordance with machine manufacturer's instructions.

COLOUR CHANGE

Depending on the product, the soda lime colour change is either white to violet or pink to white. The type of colour change will be indicated on the label, using one of the symbols indicated in the table on page 4.

STORAGE AND DISPOSAL

1. STORAGE AND HANDLING

- Store product in an environment that minimises exposure to UV light.
- Ensure adequate ventilation of the storage area. Keep containers tightly closed, at 0-35°C (32-95°F), dry and out of direct sunlight.
- Keep from freezing, which may reduce carbon dioxide absorption performance and increase dust.
- Avoid creating airborne dust, especially when pouring or handling loose material.
- Handle with care, especially when placing product in a canister.
- Soda lime carbon dioxide absorbents are classified as Corrosive and Irritant.
- Consult relevant SDS for additional information.

2. DISPOSAL

- The user shall dispose of used Soda lime carbon dioxide absorbents in compliance with all applicable local laws and regulations. Disposal may vary with the chemical or biological agents used with the absorbent (e.g. anaesthetic agents, acid-base vapours, etc.).
- Soda lime carbon dioxide absorbent that has been used with flammable anaesthetic must be kept away from heat, sparks and open flames, as residue may be present.
- See relevant SDS for Soda lime products for hazards and precautionary statements.

FEEDBACK AND SERIOUS INCIDENTS

If any serious incident has happened where this device is suspected to be a contributory factor, then this should be reported to Molecular Products Limited, using complaints@molprod.com and the competent authority in the country the device was used in.

ADDITIONAL INFO

If additional information is required please use the contact section on page 3 to request additional copies free of charge or alternatively visit Molecular Products website to download an eUser version.



Consult instructions for use



The instructions for use must be read



Indicates that the content of the device is corrosive



Do not use if package is damaged



Caution



Indicates that the contents of the device are an irritant



Expiry date



Authorized representative in the European Community/ European Union



Humidity limits



Manufacturer



Single use



UK Conformity assessment mark



Manufacture date



This way up



Fragile, handle with care



Catalogue



Unique device identifier



Temperature limit



Batch code



Indicates that the product is a medical device



Distributor



Keep away from sunlight



European CE Mark



Quantity

UFI

Indicates the unique formula number of the product



Keep dry



Period after opening

Indicating the product changes colour from pink to white after absorbing carbon dioxide

Indicating the product changes colour from white to violet after absorbing carbon dioxide

CONTACT



Molecular Products Ltd
Parkway, Harlow Business Park
Harlow, Essex, CM19 5FR, UK
T: +44 (0)1279 445111
E: sales@molprod.com
W: www.molecularproducts.com



Medical Device Management Ltd
Block B, The Crescent Building,
Northwood, Santry, Dublin 9, D09 C6X8, Ireland

Molecular Products and its subsidiaries cannot be held responsible for any damage or injury occurring, as a result of improper storage or use of its products.
Registered office as above. Reg No. 02721125 England. A member of Molecular Products Group Ltd.

